

Automatic Electronic Defibrillator Contract

	A	B	C
1	FEATURE:		PRODUCT SPECIFICATION :
2	Unit Capabilities		Full-rescue automatic electronic defibrillator with integral cardiopulmonary resuscitation assistance including monitoring/prompting of compression rate
3			
4			
5	Maximum Time Between Shocks (with fresh batteries)		Less than 20 seconds
6			
7	User Interface / Monitoring		LED or LCD ECG Display with Voice Prompt
8			
9	Battery Type		Lithium Manganese
10			
11	Battery Life		Minimum four (4) year full-operational shelf life (included with purchase)
12			
13	Battery Condition Annunciation		Continuous monitoring of battery condition, with audible beep at 1 - minute interval to indicate low battery capacity when unit is off; audible and visual warning at power on if battery condition is low or depleted
14			
15	Waveform		BiPhasic
16			
17	Energy Levels		Escalating, minimum 3-step, with Adult and Pediatric Modes
18			
19	Self - Test		4 - step: Battery Installation, Power On, Manual on demand, Periodic Automatic at programmable interval of 1 to 7 days
20			
21	Electrodes (Pads)		Included with purchase
22			
23	Pediatric Pads		Included with purchase
24			
25	Data Management		Store minimum 60 minutes clinical event data. Each unit to include free download of data control software with free upgrades through manufacturer's warranty period.
26			
27	Product Support		Toll-free number for operator questions and technical assistance
28			
29	Delivery		Ground shipping included to any ALDOT address
30			
31	Installation and Set-up		Demonstration and Set Up Videos on CD
32			

Automatic Electronic Defibrillator Contract

	A	B	C
1	FEATURE:		PRODUCT SPECIFICATION :
33	Operator Training		Each unit to include printed and electronic training materials. Each purchase to include on-line training for a minimum of one operator. Manufacturer to offer additional on-line training or in-house training as a purchase option.
34			
35	Accessories included in purchase		Adult and pediatric pads; batteries; carry case; razor, face shield, disposable gloves, wipes; AED inspection tag; wall-mounted storage cabinet with Alarm
36			
37	Documentation		Each unit to include a printed Operator's Guide and complete Administrator's Guide in either printed, electronic or web-based format.
38			
39	Physician's perscription		Each unit shall include a location - specific Certificate of Authority / Physician's Prescription for Medical Device
40			
41	Manufacturer's Warranty		Minimum five years from date of purchase
42			
43	Indemnification Agreement		Manufacturer shall indemnify, defend (or at its option, settle) and hold purchaser harmless from any and all claims, losses or expenses related to a claim of personal injury, including reasonable cost of defense, against purchaser that arise solely from the mechanical or electrical failure or malfunction of the AED device
44			